



Australian Government

Department of Health
Therapeutic Goods Administration

Australian Register of Therapeutic Goods Certificate

Issued to

AMBU Australia Pty Ltd

for approval to supply

AMBU Australia Pty Ltd - Resuscitator, pulmonary, manual, reusable

ARTG Identifier	141015
ARTG Start date	26/06/2007
Product Category	Medical Device Included Class IIa
GMDN	17591
GMDN Term	Resuscitator, pulmonary, manual, reusable
Intended Purpose	Reusable resuscitator intended for manual pulmonary resuscitation and emergency respiratory support for neonates, infants, children and adults.

Manufacturer Details	Address	Certificate number(s)
Ambu AS	Baltorpbakken 13 Ballerup, , DK-2750 Denmark	DV-2009-MC-13176-3

ARTG Standard Conditions

The above Medical Device Included Class IIa has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Resuscitator, pulmonary, manual, reusable

Product Specific Conditions

- No specific conditions have been recorded against this entry.

Therapeutic Goods Administration
PO Box 100, Woden ACT 2606 Australia
Phone: 1800 020 653
Email: info@tga.gov.au

ARTG Identifier: 141015
ARTG Start Date: 26/06/2007